



Silikontorkare för automatisk rengöringsmodul TU5300sc och TU5400sc

Produktnr:

LZY915

SEK Pris:

Kontakta oss

Tillgänglig

Maintain a clean vial with the replacement silicone wiper for the Automatic Cleaning Module. The wiper is recommended for replacement every six months, or as needed depending on sample type. For samples containing high levels of iron and manganese, order part number LZQ176.

LZY915 is the replacement item for LZQ165.