

Digital titratorpatron, natriumtiosulfat, stabiliserad, 0,02256 N

Produktnr: **2409101**

SEK Pris: Kontakta oss
Inget leveransdatum anges

For determination of Chlorine by Digital Titrator titration. 0.02256 N Digits x 0.01 = mg/L CI2 in 100 mL, Digits x 0.0000282 = meq. Stabilized sodium thiosulfate formulation.

Accurate results - when and where you need them
Digital Titrator methods use a high-precision dispensation device and cartridges of concentrated titrant to allow analyses in the lab, the plant or the field.

Fast and Flexible
Interchangeable cartridges allow you to change titrants in seconds. No cleaning or rinsing burets.

Versatile
More than 50 parameter/range combinations are available

Economical
Concentrated titrants allow 40-100 typical titrations per cartridge.

Specifikationer

Antal tester :	40 - 100
Mätområde:	1 - 400 mg/L Cl_2
Metodnamn:	Iodometric
Metodnummer:	8209
Parameter:	Chlorine, Total
Plattform:	Digital Titrator
Titrant:	0.02256 N $Na_2S_2O_3$